## Corning Incorporated Life Sciences

Registered ISO 9001:2008

## **Product Description**

Catalog Number: 5580-2

PYREX® Flask, Volumetric, Class A, 2 mL **Product Description:** 

without Snap Cap

**Component Materials:** 

Type I Class A - Low-Expansion Borosilicate Glass, per ASTM E438 (Standard Specification for Glasses in Laboratory Apparatus)

Attributes:

- E237 (Specification for Microvolumetric Vessels - Volumetric Flasks and Centrifuge Tubes) **ASTM Reference** 

E694 (Specification for Volumetric Ware) E542 (Specification for Calibration of Laboratory Volumetric Apparatus)

Capacity 2 mL ±0.015 mL

Product Dimensions - 10 mm OD x 85 mm High

Autoclaving:

This product will withstand repeated standard autoclave cycles of 121°C at 15 psi for 15 minutes.

Heating / Cooling:

This product is designed for volumetric measurement and therefore typical use will be at standard room temperature. Volumetric glassware should not be subjected to temperatures in excess of 150°C.

Vacuum / Pressure:

This product is not designed for and is not rated for vacuum or pressure applications.

Sterilization:

Shipped Not Sterile

Lot Number Designation:

Each production lot is uniquely identified with an 8 digit lot number as follows: First 3 digits - Julian date, start of manufacturing; Next 2 digits - Year of manufacture; Last 3 digits - Batch identification.

**Revision Date:** 

1 AUG 2013

Rev No: 1