

**Corning Incorporated  
Life Sciences**

**Registered  
ISO 9001:2008**

## Product Description

**Catalog Number:** 5340-1L

**Product Description:** PYREX® Flask, Filtering, Heavy Wall, Graduated, Tubulated, 1000 mL  
Tooled for Rubber Stopper #8

**Component Materials:**

Type I Class A - Low-Expansion Borosilicate Glass, per ASTM E438 (Standard Specification for Glasses in Laboratory Apparatus)

**Attributes:**

Capacity - 1000 mL ±50 mL at any mark (5% of full capacity)  
Product Dimensions - 137 mm OD x 238 mm High  
Tubulation - 10 mm OD

**Autoclaving:**

This product will withstand repeated standard autoclave cycles of 121°C at 15 psi for 15 minutes.

**Heating / Cooling:**

This product should NOT be heated while being used in a vacuum or pressure application as the combination of thermal and mechanical stresses could cause a failure.

This product is not designed for freezing applications.

**Vacuum / Pressure:**

This product is designed for full vacuum of 760 mm Mercury (15 psi negative pressure).

This product is not designed for and is not rated for pressure applications.

**Sterilization:**

Shipped Not Sterile

**TSE/BSE:**

Materials of animal origin are not used in the manufacture of this product.

**Lot Number Designation:**

Each production lot is uniquely identified with an 8 digit lot number as follows: First 3 digits - Julian date, start of manufacturing; Next 2 digits - Year of manufacture; Last 3 digits - Batch identification.

**Revision Date:**

19 JUL 2013

Rev No: 1